

## Damaged Sample Cups for Flashpoint Analyzer

### Note:

The Sample Cup must never be damaged or dented (if dropped). A damaged or dented cup will result in Erroneous Test Results.

### How to remove minor scratches:

Minor dings and scratches can be removed by lightly sanding with very fine sandpaper (400-600),

If the nickel plating is missing or pitting, please replace the sample cup.

Both versions are available in SS.

